

**PSA (Clone: APSP)
Rabbit Monoclonal Antibody**

PRODUCT INFORMATION:

REF	
MR1347	6ml Ready to use
MR1347	3ml Ready to use
MRC1347	1ml Concentrated
MRC1347	0.5ml Concentrated
MRC1347	0.1ml Concentrated
MRH1347	6ml Ready to use
MRH1347	3ml Ready to use

PERFORMANCE CHARACTERISTICS:

Localization: Cytoplasm
Retrieval Buffer: Tris-EDTA, pH 9.0
Incubation: 30-60 minutes
Positive control: Prostate carcinoma,
 Benign Prostate
 Hyperplasia

INTENDED USE

For research use only

This antibody is intended for use in qualitatively identifying CD23 antigen by light microscopy in formalin-fixed, paraffin-embedded (FFPE) tissue sections using immunohistochemical (IHC) detection methodology. Any observed staining or its absence should be reviewed alongside appropriate positive and negative controls. Interpretation must be carried out by qualified personnel trained in histological and molecular techniques. This product is for research use only and not intended for diagnostic or therapeutic use.

SUMMARY AND EXPLANATION

Prostate-specific antigen (PSA) is a single-chain glycoprotein of 237 amino acids containing approximately 8% carbohydrate. It is a serine protease produced almost exclusively by prostatic epithelial cells. Immunohistochemically PSA is expressed in the highly specialized apically-superficial layer of female and male secretory cells of the prostate gland, and is readily demonstrated in adenocarcinomas of the prostate in about 99% of the cases. There is a correlation between malignancy grade and intensity of staining, high grade carcinomas displaying weaker expression. About 1% of poorly differentiated carcinomas have been negative for PSA. Due to the high specificity of PSA for prostatic glandular epithelium, it is very useful in identifying prostatic carcinoma in the prostate and in the adjacent organs often affected by epithelial malignancies, i.e. rectum and urinary bladder. PSA may be used in a panel together with NKX3.1 and Prostein, which are at least as sensitive and slightly more specific than PSA. PSA also known as KLK3, APS, KLK2A1, PSA, hK3, kallikrein related peptidase 3, Prostate Specific Antigen.

PRINCIPLE OF THE PROCEDURE

The identification of the antigen on the FFPE tissues is carried out using the above-stated antibody. The antigen and antibody complex are visualized using an enzyme-coupled (HRP/AP) secondary antibody with specific binding to the primary antibody, this complex is visualized by the enzymatic activation of the chromogen resulting in a visible reaction production of the antigenic site. Every step involves precise time and optimal temperature and the results are interpreted using a light microscope by a qualified and trained pathologist.

REAGENT PROVIDED

Concentrated format: PSA is affinity purified and diluted in antibody diluent with 1% bovine serum albumin (BSA) and 0.05% of sodium azide (NaN₃).
Recommended dilutions: 1:50 – 1:100
 The antibody dilution and protocol may vary depending on the specimen preparation and specific application. Optimal conditions should be determined by individual laboratories.
Pre-diluted format: PathnSitu's ready-to-use antibodies are pre-titrated to optimal staining conditions. Further dilution will affect the efficacy of the antibody and may yield to sub-optimal staining.
Immunogen: Synthetic peptide corresponding to PSA residues within aa150-250 of PSA was used as an immunogen.
Host, Isotype: Rabbit, IgG

STORAGE AND HANDLING

Storage Recommendations: Store at 2-8°C. When stored at the appropriate conditions, the antibody is stable until expiry. Do not use the antibody after the expiration date provided on the vial in any condition. To ensure proper reagent delivery and stability, replace the dispenser cap after

SPECIMEN PREPARATION

Staining Recommendations:

Routinely processed, FFPE tissues are suitable for use with this primary antibody, when using PathnSitu's Poly Excel HRP/DAB detection system. The recommended tissue fixative is 10% neutral buffered formalin. Variable results may occur as a result of prolonged fixation or special processes such as decalcification. The thickness of the sections should be 2-5µm. Slides should be stained once the sections are made as the cut sections' antigenicity may diminish over time. Staining known positive and negative controls simultaneously with unknown specimens is recommended.

PRECAUTIONS

1. This product should be used by qualified and trained professional users only
2. The product contains < 0.1% of sodium azide as a preservative and is not classified as hazardous, refer to MSDS for further details
3. As with any product derived from biological sources, proper handling procedures should be used
4. Do not use reagents after the expiration date
5. Use protective clothing and gloves, while handling reagents
6. All hazardous materials should be disposed of according to local state and federal regulations
7. Avoid microbial contamination of reagents as it may lead to incorrect results

STAINING PROCEDURE

Antigen Retrieval Solution: Use Tris-EDTA buffer (Cat# PS009) as an antigen retrieval solution.
Heat Retrieval Method: Retrieve sections under steam pressure for 15 minutes using PathnSitu's MERS (Multi Epitope Retrieval System) for optimal retrieval of the epitopes, allow solution to cool at room temperature, transfer the tissue sections/slides to the distilled water before the primary antibody application.
Primary Antibody: Cover the tissue sections with primary antibody and incubate for 30-60 min at room temperature when used PathnSitu's PolyExcel Detection System.
Detection System: Refer to PathnSitu's PolyExcel HRP/ DAB detection system protocol for optimal staining results.

QUALITY CONTROL

The recommended positive tissue controls for PSA are Prostate carcinoma and Benign Prostate hyperplasia. A positive and negative tissue controls must be run with every staining procedure performed to monitor the correct performance of processed tissue and test reagents. A negative tissue control provides an indication of non-specific background staining. If the results are not expected in positive and negative controls the test must be considered invalid and the entire procedure must be cross-verified. The individual laboratory must establish their own quality control to validate the process and antibody when opening a vial.

INTERPRETATION OF RESULTS

PSA stains the Cytoplasm. A qualified experienced/trained pathologist must interpret the results in the patient's sample along with the positive and negative controls.

ANALYTIC PERFORMANCE CHARACTERISTICS

1. Heat the paraffin-embedded tissue slides for a suitable duration at an appropriate temperature to promote tissue adhesion.
 Note: Use positively charged coated slides (Cat no.: PS-011-72) for better adherence.
2. Deparaffinize the slides using xylene (preferably 3 changes with 5min each) to clear the paraffin wax present on and around the tissue.
3. Rehydrate the slides in graded alcohols (100%, 70%, and 50%) for 3 min each and in distilled water (preferably 2 changes with 2 min each) respectively.
4. Immerse the slides in 1X retrieval buffer (preferable Cat No.: PS009) and subject them to Heat-induced epitope retrieval by using a multi-epitope retrieval system (MERS-i) to unmask the epitopes.

5. Proceed further by using Poly Excel DAB Detection system (preferably Cat no: PEH002 or OSH001) kit components like Poly Excel Peroxidase Block to inactivate or block the non-specific binding firstly.
6. Apply the primary antibody specific to the target antigen. Incubate slides with the primary antibody for a suitable duration at an appropriate temperature as mentioned in the datasheet.
7. Rinse the slides to remove unbound primary antibody using wash buffer (preferably Cat no: PS006)
8. Apply the secondary antibody (preferably Poly Excel Poly HRP- Anti-Mouse/Anti-Rabbit Cat no: PEH002 or OSH001) conjugated to an enzyme that recognizes the primary antibody. Incubate slides with the secondary antibody for a suitable duration at an appropriate temperature.
9. Rinse the slides to remove unbound secondary antibodies using wash buffer (Preferably Immunowash buffer Cat no: PS006)
10. Apply a substrate, PolyExcel Stunn DAB Chromogen for enzyme-conjugated secondary antibody for a suitable duration.
11. Counter-stain the tissue section to visualize the expression in specific structures or cell types.
12. Dehydrate slides through graded alcohols (70%,90%, 100%,100%), clear the slides in Xylene (preferably 3 changes with 2min each) and mount the slides with an appropriate mounting medium.
13. Visualize the stained slides under the microscope.

The antibody consistently exhibited specific and sensitive staining across various positive and negative tissue controls, including Prostate Carcinoma, BPH and SCC of Tongue tissue samples with cytoplasm staining. This specificity and sensitivity were validated through inter-run, intra-run, and lot-based studies. The stability of the antibody which was determined using real-time or accelerated methods extends until the expiration date indicated on the product labels.

TROUBLESHOOTING

1. Follow the antibody-specific protocol recommendations according to the datasheet provided
2. Tissue staining is dependent on the handling and processing of the tissue prior to staining. Improper fixation, tissue processing, antibody freezing and thawing, washing, drying, heating, sectioning or contamination with other tissues or fluids may produce artifacts, antibody trapping or inaccurate results
3. Do not allow the section to dry out during the entire IHC process
4. Excessive or incomplete counterstaining may compromise the interpretation of the results
5. If unusual results occur, contact PathnSitu's Technical Support at +91-40-2701 5544 or E-mail: techsupport@pathnsitu.com

LIMITATIONS AND WARRANTY

1. Authorized and skilled/trained personnel only may use the product.
2. The clinical interpretation of any test results should be evaluated within the context of the patient's medical history and other diagnostic test results.
3. A qualified trained pathologist must perform the evaluation of the test results.
4. The product comes with no warranties beyond the provided description
5. Use appropriate volume/concentration to cover entire tissue sections and optimum conditions to avoid false positive and negative results.
6. Use appropriate/recommended buffer/instruments/all consumables with appropriate incubation timings to obtain optimal results.
7. Always recommend using known positive and negative controls to evaluate the test result.
8. Unexpected reactions may occur in untested tissues due to tissue component variability.
9. False positive results can arise from no stringent washing practices and other contributing factors.
10. In instances where localization differs from the specifications outlined in the datasheet, clinical coordination or prompt technical support is advised.
11. Maintain recommended storage conditions.
12. Refer entire data sheet to know any further limitations about the product.
13. No warranties whether expressed or implied, extend beyond the description.
14. PathnSitu is not liable for property damage, personal injury, time or effort or economic loss caused by this product.

BIBLIOGRAPHY

1. Ward AM, et.al, Ann Clin Biochem 2001; 38: 633-651.
2. ACS Aguiar , et.al, Histol Histopathol (2000) 15: 131-142

EXPLANATION OF SYMBOLS

LOT

Lot number / Batch number



Expiry

RUO

Research use only



Storage limitation



Date of manufacture

REF

Catalogue number



Manufacturer address